

Radon Detector Instructions

Turn the detector over and push the reset button with a pencil.

The device will go through a “CAL” calibration mode. This will take a few hours.

Set the unit in a desired location. Residential basement or cottage lowest floor are preferred but any area that is lived in is acceptable.

Leave the unit sitting untouched for 6 days. Set your calendar to remind you that the detector needs to be returned so the next member can pick it up to test their residence.

On Day 6, read the LONG-TERM AVERAGE number.

The SHORT-TERM AVERAGE is a daily count and is calculated into the long-term average.

The World Health Organization recommends that the average concentration be below 100Bq/m³. If your reading is above 100, contact a professional to perform an additional test on the residence.

Return the detector to the storage box located where you picked it up.

Lock Box code is (951) 1163 Maplehurst Dr. on the side wall of container.

Further information can be located on the PLA web site under the Environmental tab.

The PLA is not responsible for any reading of concern nor do they offer any professional guidance or assistance. This is an informational tool only to the PLA and PLCA membership.